



This Quick Start Reference provides information on setting up your instrument for basic operation. The latest complete Communication and Operational Manual as well as free Software and ActiveX Controls are available at [www.omega.com/specs/iseries](http://www.omega.com/specs/iseries).

SAFETY CONSIDERATION

**This device is marked with the international Caution symbol.**

The instrument is a panel mount device protected in accordance with 2014/35/EU. Remember that the unit has no power-on switch. Building installation should include a switch or circuit-breaker that must be compliant to IEC 947-1 and 947-3.


SAFETY:

- Do not exceed voltage rating on the label located on the top of the instrument housing.
- Always disconnect power before changing signal and power connections.
- Do not use this instrument on a work bench without its case for safety reason.
- Do not operate this instrument in flammable or explosive atmospheres.
- Do not expose this instrument to rain or moisture.

EMC:

- Whenever EMC is an issue, always use shielded cables.
- Never run signal and power wires in the same conduit.
- Use signal wire connections with twisted-pair cables.
- Install Ferrite Bead(s) on signal wire close to the instrument if EMC problems persist.

MOUNTING



Panel Mounting Instruction:

- Using the dimensions from the panel cutout diagram shown above, cut an opening in the panel.
- Insert the unit into the opening from the front of the panel, so the gasket seals between the bezel and the front of the panel.
- Slide the retainer over the rear of the case and tighten against the backside of the mounting panel.

Disassembly instruction:

If necessary, the unit may be removed from the panel and opened.


**Warning: Disconnect all ac power from the unit before proceeding.**

- Make sure the AC power is disconnected.
- Remove all wiring connections from the rear of the meter. To remove power and input connectors squeeze top and bottom of the case near the connector site for release, then pull connectors from the meter.
- To remove meter from the case, squeeze top and bottom of the bezel to release, then pull from case.


WIRING

Wire the instrument according to the figure shown below.

**Warning: Do not connect ac power to your device until you have completed all input and output connections. This device must only be installed by a specially trained electrician with corresponding qualifications. Failure to follow all instructions and warnings may result in injury!**



Connect the main power connections as shown in the figure below.



FUSE	Connector	Output Type	For 115Vac	For 230Vac	DC
FUSE 1	Output 1	Relay	3 A(T)	3 A(T)	-
FUSE 2	Output 2	Relay	3 A(T)	3 A(T)	-
FUSE 3	Power	N/A	100 mA(T)	100 mA(T)	100 mA(T)
FUSE 4	Power	N/A	N/A	N/A	400 mA(T)

**Note** Output 1 and 2 are for -AL Limit Alarm Option only.

CONFIGURATION

MENU Mode:

Flashing display in MENU Mode means you can make your selection by pressing **➡** button. If the flashing display is not a four digit value, pressing **➡** button will always direct the instrument one step backward of the top menu item. The second push on the **➡** button will reset the instrument except after the setpoint and the alarms, that will go to the RUN Mode without resetting the instrument. The **⬅** button will always sequence the instrument thru the menu items.


The **⬅** button has two functions:

- To save a selected flashing display
- To direct the instrument to the next submenu level

RUN Mode:

- ➡** causes the display to flash the PEAK with the corresponding value. Press again to go back to RUN Mode.
- ⬅** causes the display to flash VALLEY with the corresponding value. Press again to go back to RUN Mode.
- ⬅** causes flashing PEAK or VALLEY to reset corresponding values. Press **⬅** one more time to go back to RUN Mode.

FLOW CHART



Underline denotes factory default setup

OPERATION - (For Thermocouple Input)

Step 1. Apply Power to the Instrument

When your device is first powered up it will display the ambient temperature (assume 75°F).

Step 2. Enter Setpoint 1 Menu

Press **⬅** one time from run mode to get to **SP.1** Setpoint 1.

Step 3. Enter the Setpoint 1 Value Submenu

Press **⬅**. Display shows the previous selection of Setpoint 1.

Step 4. Change the Setpoint 1 Value

Press **➡** or **⬅** until desired value is displayed.

Step 5. Store the Setpoint 1 Value

Set the Setpoint 1 to 10 degree higher than Process value (SP1 = 85) and press **⬅** to store, display flashes **SE.P.1** message and advances to **SP.2** Setpoint 2 Menu.

Step 6. Store the Setpoint 2 value

Repeat steps 3 and 4. Set the Setpoint 2 to 5 degree higher than Process value (SP2 = 80) and press **⬅** to store, display flashes **SE.P.2** message and advances to **CNFG** Configuration Menu.

Step 7. Enter the Input Type Menu

Press **⬅** to enter **INPT** Input Type Menu.

Step 8. Enter to the submenu items of Input Menu

Press **⬅** to display Input: Process, RTD or Thermocouple. If flashing **E.C** is displayed press **⬅** and proceed to step 11.

Step 9. Scroll through available selection of Input Menu  
Press **➡** until a flashing **E.C** for Thermocouple is displayed.